

# Notice of Allowability

Application No.

10/690,868

Examiner

Feben M. Haile

Applicant(s)

GOODE ET AL.

Art Unit

2616

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to application filed October 22, 2003.
2. ☒ The allowed claim(s) is/are 1-8.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☐ All b) ☐ Some\* c) ☐ None of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftperson's Patent Drawing Review (PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

## Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date 10/22/2003
4. ☐ Examiner's Comment Regarding Requirement for Deposit  
of Biological Material

5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_.
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_.



DORIS H. TO  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 2600

## **DETAILED ACTION**

### ***Allowable Subject Matter***

1. Claims 1-8 allowed. The following is an examiner's statement of reasons for allowance:

In conventional schemes, digital information can be distributed through networks such as Asynchronous Transfer Mode and SONET. However both these systems are not cost effective and require significant overhead bandwidth. A more efficient network solution for the delivery of digital data is a European Standard known as Digital Video Asynchronous Serial Interface. This specification defines a protocol for distributing the digital data at an actual data carriage of 216 Mbps. Although this is fine for point-to-point communications, it does not utilize bandwidth efficiently and support high-speed data transmissions necessary for multiple users.

In the instant application, the disadvantages associated with the prior art are overcome by a facilitating a multiple user digital information distribution arrangement having a counter-rotating Asynchronous Serial Interface ring network. This system enables delivery of digital information through a topology comprising a plurality of nodes. Within each node, a ring terminal extracts packets addressed to that particular node and inserts a character for synchronization purposes. This scheme provides high-speed transmissions with low overheads and an efficient use of bandwidth.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

**Conclusion**


2. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure:

a) Mitchell (US 6,442,134), ATM Network with Ring Architecture

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Feben M. Haile whose telephone number is (571) 272-3072. The examiner can normally be reached on 6:00am - 3:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Doris To can be reached on (571) 272-7629. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

  
0413012007